



December 02, 2025

Lauren Doherty
Intertek PSI
17 British American Boulevard
Latham, NY 12110

RE: Project: BURNT HILLS MS 11/6
Pace Project No.: 70391829

Dear Lauren Doherty:

Enclosed are the analytical results for sample(s) received by the laboratory on November 15, 2025. The results relate only to the samples included in this report. Results reported herein conform to the applicable TNI/NELAC Standards and the laboratory's Quality Manual, where applicable, unless otherwise noted in the body of the report.

The test results provided in this final report were generated by each of the following laboratories within the Pace Network:

- Pace Analytical Services - Melville

If you have any questions concerning this report, please feel free to contact me.

Sincerely,

A handwritten signature in cursive script that reads "Lori Beyer".

Lori A. Beyer
lori.beyer@pacelabs.com
516-370-6014
Project Manager

Enclosures

cc: Lauren Doherty, Intertek PSI



REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



CERTIFICATIONS

Project: BURNT HILLS MS 11/6

Pace Project No.: 70391829

Pace Analytical Services, LLC - Melville, NY

575 Broad Hollow Rd, Melville, NY 11747

Connecticut Certification #: PH-0435

Delaware Certification # NY 10478

Maryland Certification #: 208

Massachusetts Certification #: M-NY026

New Hampshire Certification #: 2987

New Jersey Certification #: NY158

New York Certification #: 10478 Primary Accrediting Body

Pennsylvania Certification #: 68-00350

Rhode Island Certification #: LAO00340

Texas Certification #: T104704582

Florida Certification #: E871198

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



SAMPLE SUMMARY

Project: BURNT HILLS MS 11/6

Pace Project No.: 70391829

Lab ID	Sample ID	Matrix	Date Collected	Date Received
70391829001	MS-002	Drinking Water	11/06/25 04:47	11/15/25 07:37
70391829002	MS-004	Drinking Water	11/06/25 04:49	11/15/25 07:37
70391829003	MS-009	Drinking Water	11/06/25 04:52	11/15/25 07:37
70391829004	MS-010	Drinking Water	11/06/25 04:54	11/15/25 07:37
70391829005	MS-024	Drinking Water	11/06/25 05:01	11/15/25 07:37
70391829006	MS-025	Drinking Water	11/06/25 05:03	11/15/25 07:37
70391829007	MS-058	Drinking Water	11/06/25 05:08	11/15/25 07:37
70391829008	MS-059	Drinking Water	11/06/25 05:08	11/15/25 07:37
70391829009	MS-060	Drinking Water	11/06/25 05:08	11/15/25 07:37
70391829010	MS-061	Drinking Water	11/06/25 05:09	11/15/25 07:37
70391829011	MS-062	Drinking Water	11/06/25 05:10	11/15/25 07:37
70391829012	MS-063	Drinking Water	11/06/25 05:10	11/15/25 07:37
70391829013	MS-072	Drinking Water	11/06/25 05:13	11/15/25 07:37
70391829014	MS-073	Drinking Water	11/06/25 05:14	11/15/25 07:37
70391829015	MS-080	Drinking Water	11/06/25 05:17	11/15/25 07:37
70391829016	MS-081	Drinking Water	11/06/25 05:18	11/15/25 07:37
70391829017	MS-084	Drinking Water	11/06/25 05:20	11/15/25 07:37
70391829018	MS-085	Drinking Water	11/06/25 05:35	11/15/25 07:37
70391829019	MS-086	Drinking Water	11/06/25 05:36	11/15/25 07:37
70391829020	MS-088	Drinking Water	11/06/25 05:41	11/15/25 07:37
70391829021	MS-089	Drinking Water	11/06/25 05:41	11/15/25 07:37
70391829022	MS-104	Drinking Water	11/06/25 05:39	11/15/25 07:37
70391829023	MS-105	Drinking Water	11/06/25 05:44	11/15/25 07:37
70391829024	MS-107	Drinking Water	11/06/25 05:32	11/15/25 07:37
70391829025	MS-151	Drinking Water	11/06/25 05:28	11/15/25 07:37
70391829026	MS-153	Drinking Water	11/06/25 05:21	11/15/25 07:37
70391829027	MS-154	Drinking Water	11/06/25 05:22	11/15/25 07:37
70391829028	MS-155	Drinking Water	11/06/25 05:25	11/15/25 07:37
70391829029	MS-156	Drinking Water	11/06/25 04:57	11/15/25 07:37
70391829030	MS-157	Drinking Water	11/06/25 04:58	11/15/25 07:37
70391829031	MS-158	Drinking Water	11/06/25 04:59	11/15/25 07:37
70391829032	MS-159	Drinking Water	11/06/25 05:48	11/15/25 07:37
70391829033	MS-160	Drinking Water	11/06/25 05:49	11/15/25 07:37

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



SAMPLE ANALYTE COUNT

Project: BURNT HILLS MS 11/6

Pace Project No.: 70391829

Lab ID	Sample ID	Method	Analysts	Analytes Reported
70391829001	MS-002	EPA 200.8	JP2	1
70391829002	MS-004	EPA 200.8	JP2	1
70391829003	MS-009	EPA 200.8	JP2	1
70391829004	MS-010	EPA 200.8	JP2	1
70391829005	MS-024	EPA 200.8	JP2	1
70391829006	MS-025	EPA 200.8	JP2	1
70391829007	MS-058	EPA 200.8	JP2	1
70391829008	MS-059	EPA 200.8	JP2	1
70391829009	MS-060	EPA 200.8	JP2	1
70391829010	MS-061	EPA 200.8	JP2	1
70391829011	MS-062	EPA 200.8	JP2	1
70391829012	MS-063	EPA 200.8	JP2	1
70391829013	MS-072	EPA 200.8	JP2	1
70391829014	MS-073	EPA 200.8	JP2	1
70391829015	MS-080	EPA 200.8	JP2	1
70391829016	MS-081	EPA 200.8	JP2	1
70391829017	MS-084	EPA 200.8	JP2	1
70391829018	MS-085	EPA 200.8	JP2	1
70391829019	MS-086	EPA 200.8	JP2	1
70391829020	MS-088	EPA 200.8	JP2	1
70391829021	MS-089	EPA 200.8	JP2	1
70391829022	MS-104	EPA 200.8	JP2	1
70391829023	MS-105	EPA 200.8	JP2	1
70391829024	MS-107	EPA 200.8	JP2	1
70391829025	MS-151	EPA 200.8	JP2	1
70391829026	MS-153	EPA 200.8	JP2	1
70391829027	MS-154	EPA 200.8	JP2	1
70391829028	MS-155	EPA 200.8	JP2	1
70391829029	MS-156	EPA 200.8	JP2	1
70391829030	MS-157	EPA 200.8	JP2	1
70391829031	MS-158	EPA 200.8	JP2	1
70391829032	MS-159	EPA 200.8	JP2	1
70391829033	MS-160	EPA 200.8	JP2	1

PACE-MV = Pace Analytical Services - Melville

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: BURNT HILLS MS 11/6

Pace Project No.: 70391829

Sample: MS-002		Lab ID: 70391829001		Collected: 11/06/25 04:47		Received: 11/15/25 07:37		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	2.7	ug/L	1.0	1		12/01/25 17:03	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: BURNT HILLS MS 11/6

Pace Project No.: 70391829

Sample: MS-004		Lab ID: 70391829002	Collected: 11/06/25 04:49	Received: 11/15/25 07:37	Matrix: Drinking Water			
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville						
Lead	<1.0	ug/L	1.0	1		12/01/25 17:04	7439-92-1	

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: BURNT HILLS MS 11/6

Pace Project No.: 70391829

Sample: MS-009		Lab ID: 70391829003		Collected: 11/06/25 04:52		Received: 11/15/25 07:37		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	1.8	ug/L	1.0	1		12/01/25 17:06	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: BURNT HILLS MS 11/6

Pace Project No.: 70391829

Sample: MS-010		Lab ID: 70391829004		Collected: 11/06/25 04:54		Received: 11/15/25 07:37		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	1.4	ug/L	1.0	1		12/01/25 17:07	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: BURNT HILLS MS 11/6

Pace Project No.: 70391829

Sample: MS-024		Lab ID: 70391829005		Collected: 11/06/25 05:01		Received: 11/15/25 07:37		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		12/01/25 17:09	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: BURNT HILLS MS 11/6

Pace Project No.: 70391829

Sample: MS-025		Lab ID: 70391829006		Collected: 11/06/25 05:03		Received: 11/15/25 07:37		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		12/01/25 17:10	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: BURNT HILLS MS 11/6

Pace Project No.: 70391829

Sample: MS-058		Lab ID: 70391829007		Collected: 11/06/25 05:08		Received: 11/15/25 07:37		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	1.7	ug/L	1.0	1		12/01/25 17:15	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: BURNT HILLS MS 11/6

Pace Project No.: 70391829

Sample: MS-059		Lab ID: 70391829008		Collected: 11/06/25 05:08		Received: 11/15/25 07:37		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	1.4	ug/L	1.0	1		12/01/25 17:22	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: BURNT HILLS MS 11/6

Pace Project No.: 70391829

Sample: MS-060		Lab ID: 70391829009		Collected: 11/06/25 05:08		Received: 11/15/25 07:37		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	1.5	ug/L	1.0	1		12/01/25 17:27	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: BURNT HILLS MS 11/6
Pace Project No.: 70391829

Sample: MS-061		Lab ID: 70391829010		Collected: 11/06/25 05:09		Received: 11/15/25 07:37		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	1.2	ug/L	1.0	1		12/01/25 17:28	7439-92-1		

REPORT OF LABORATORY ANALYSIS



ANALYTICAL RESULTS

Project: BURNT HILLS MS 11/6

Pace Project No.: 70391829

Sample: MS-062		Lab ID: 70391829011		Collected: 11/06/25 05:10		Received: 11/15/25 07:37		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	2.0	ug/L	1.0	1		12/01/25 17:30	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: BURNT HILLS MS 11/6

Pace Project No.: 70391829

Sample: MS-063		Lab ID: 70391829012		Collected: 11/06/25 05:10		Received: 11/15/25 07:37		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	1.2	ug/L	1.0	1		12/01/25 17:31	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: BURNT HILLS MS 11/6

Pace Project No.: 70391829

Sample: MS-072		Lab ID: 70391829013		Collected: 11/06/25 05:13		Received: 11/15/25 07:37		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		12/01/25 17:33	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: BURNT HILLS MS 11/6

Pace Project No.: 70391829

Sample: MS-073		Lab ID: 70391829014		Collected: 11/06/25 05:14		Received: 11/15/25 07:37		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		12/01/25 17:34	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: BURNT HILLS MS 11/6

Pace Project No.: 70391829

Sample: MS-080		Lab ID: 70391829015		Collected: 11/06/25 05:17		Received: 11/15/25 07:37		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		12/01/25 17:39	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: BURNT HILLS MS 11/6

Pace Project No.: 70391829

Sample: MS-081		Lab ID: 70391829016		Collected: 11/06/25 05:18		Received: 11/15/25 07:37		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		12/01/25 17:40	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: BURNT HILLS MS 11/6

Pace Project No.: 70391829

Sample: MS-084		Lab ID: 70391829017		Collected: 11/06/25 05:20		Received: 11/15/25 07:37		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		12/01/25 17:41	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: BURNT HILLS MS 11/6

Pace Project No.: 70391829

Sample: MS-085		Lab ID: 70391829018		Collected: 11/06/25 05:35		Received: 11/15/25 07:37		Matrix: Drinking Water	
Parameters	Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual	
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	2.7	ug/L	1.0	1		12/01/25 17:43	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: BURNT HILLS MS 11/6
Pace Project No.: 70391829

Sample: MS-086		Lab ID: 70391829019		Collected: 11/06/25 05:36		Received: 11/15/25 07:37		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	1.8	ug/L	1.0	1		12/01/25 17:44	7439-92-1		

REPORT OF LABORATORY ANALYSIS



ANALYTICAL RESULTS

Project: BURNT HILLS MS 11/6

Pace Project No.: 70391829

Sample: MS-088		Lab ID: 70391829020		Collected: 11/06/25 05:41		Received: 11/15/25 07:37		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	1.7	ug/L	1.0	1		12/01/25 17:46	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: BURNT HILLS MS 11/6
Pace Project No.: 70391829

Sample: MS-089		Lab ID: 70391829021		Collected: 11/06/25 05:41		Received: 11/15/25 07:37		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	1.7	ug/L	1.0	1		12/01/25 17:47	7439-92-1		

REPORT OF LABORATORY ANALYSIS



ANALYTICAL RESULTS

Project: BURNT HILLS MS 11/6

Pace Project No.: 70391829

Sample: MS-104		Lab ID: 70391829022		Collected: 11/06/25 05:39		Received: 11/15/25 07:37		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		12/01/25 17:49	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: BURNT HILLS MS 11/6

Pace Project No.: 70391829

Sample: MS-105		Lab ID: 70391829023		Collected: 11/06/25 05:44		Received: 11/15/25 07:37		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	2.9	ug/L	1.0	1		12/01/25 17:50	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: BURNT HILLS MS 11/6

Pace Project No.: 70391829

Sample: MS-107		Lab ID: 70391829024		Collected: 11/06/25 05:32		Received: 11/15/25 07:37		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		12/01/25 17:52	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: BURNT HILLS MS 11/6

Pace Project No.: 70391829

Sample: MS-151		Lab ID: 70391829025		Collected: 11/06/25 05:28		Received: 11/15/25 07:37		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		12/01/25 17:56	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: BURNT HILLS MS 11/6

Pace Project No.: 70391829

Sample: MS-153		Lab ID: 70391829026		Collected: 11/06/25 05:21		Received: 11/15/25 07:37		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		12/01/25 17:58	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: BURNT HILLS MS 11/6
Pace Project No.: 70391829

Sample: MS-154		Lab ID: 70391829027		Collected: 11/06/25 05:22		Received: 11/15/25 07:37		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		12/01/25 18:02	7439-92-1		

REPORT OF LABORATORY ANALYSIS



ANALYTICAL RESULTS

Project: BURNT HILLS MS 11/6

Pace Project No.: 70391829

Sample: MS-155		Lab ID: 70391829028		Collected: 11/06/25 05:25		Received: 11/15/25 07:37		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	14.8	ug/L	1.0	1		12/01/25 18:07	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: BURNT HILLS MS 11/6

Pace Project No.: 70391829

Sample: MS-156		Lab ID: 70391829029		Collected: 11/06/25 04:57		Received: 11/15/25 07:37		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	3.2	ug/L	1.0	1		12/01/25 18:14	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: BURNT HILLS MS 11/6

Pace Project No.: 70391829

Sample: MS-157		Lab ID: 70391829030		Collected: 11/06/25 04:58		Received: 11/15/25 07:37		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	4.6	ug/L	1.0	1		12/01/25 18:15	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: BURNT HILLS MS 11/6

Pace Project No.: 70391829

Sample: MS-158		Lab ID: 70391829031		Collected: 11/06/25 04:59		Received: 11/15/25 07:37		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	3.7	ug/L	1.0	1		12/01/25 18:17	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: BURNT HILLS MS 11/6

Pace Project No.: 70391829

Sample: MS-159		Lab ID: 70391829032		Collected: 11/06/25 05:48		Received: 11/15/25 07:37		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		12/01/25 18:18	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



ANALYTICAL RESULTS

Project: BURNT HILLS MS 11/6

Pace Project No.: 70391829

Sample: MS-160		Lab ID: 70391829033		Collected: 11/06/25 05:49		Received: 11/15/25 07:37		Matrix: Drinking Water	
Parameters		Results	Units	Report Limit	DF	Prepared	Analyzed	CAS No.	Qual
200.8 MET ICPMS Drinking Water		Analytical Method: EPA 200.8 Pace Analytical Services - Melville							
Lead	<1.0	ug/L	1.0	1		12/01/25 18:20	7439-92-1		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



QUALITY CONTROL DATA

Project: BURNT HILLS MS 11/6
Pace Project No.: 70391829

QC Batch:	430441	Analysis Method:	EPA 200.8
QC Batch Method:	EPA 200.8	Analysis Description:	200.8 MET No Prep Drinking Water
		Laboratory:	Pace Analytical Services - Melville

Associated Lab Samples: 70391829001, 70391829002, 70391829003, 70391829004, 70391829005, 70391829006

METHOD BLANK: 2302865 Matrix: Water
Associated Lab Samples: 70391829001, 70391829002, 70391829003, 70391829004, 70391829005, 70391829006

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Lead	ug/L	<1.0	1.0	12/01/25 16:27	

LABORATORY CONTROL SAMPLE: 2302866

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	50	49.6	99	85-115	

MATRIX SPIKE SAMPLE: 2302868

Parameter	Units	70392887002 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	<1.0	50	50.4	100	70-130	

MATRIX SPIKE SAMPLE: 2302870

Parameter	Units	70392949001 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	<1.0	50	54.0	107	70-130	

SAMPLE DUPLICATE: 2302867

Parameter	Units	70392887002 Result	Dup Result	RPD	Max RPD	Qualifiers
Lead	ug/L	<1.0	<1.0		20	

SAMPLE DUPLICATE: 2302869

Parameter	Units	70392949001 Result	Dup Result	RPD	Max RPD	Qualifiers
Lead	ug/L	<1.0	<1.0		20	

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



QUALITY CONTROL DATA

Project: BURNT HILLS MS 11/6

Pace Project No.: 70391829

QC Batch:	430488	Analysis Method:	EPA 200.8
QC Batch Method:	EPA 200.8	Analysis Description:	200.8 MET No Prep Drinking Water
		Laboratory:	Pace Analytical Services - Melville
Associated Lab Samples:	70391829007, 70391829008, 70391829009, 70391829010, 70391829011, 70391829012, 70391829013, 70391829014, 70391829015, 70391829016, 70391829017, 70391829018, 70391829019, 70391829020, 70391829021, 70391829022, 70391829023, 70391829024, 70391829025, 70391829026		

METHOD BLANK:	2303223	Matrix:	Water
Associated Lab Samples:	70391829007, 70391829008, 70391829009, 70391829010, 70391829011, 70391829012, 70391829013, 70391829014, 70391829015, 70391829016, 70391829017, 70391829018, 70391829019, 70391829020, 70391829021, 70391829022, 70391829023, 70391829024, 70391829025, 70391829026		

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Lead	ug/L	<1.0	1.0	12/01/25 17:12	

LABORATORY CONTROL SAMPLE:	2303224					
Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	50	48.8	98	85-115	

MATRIX SPIKE SAMPLE:	2303226						
Parameter	Units	70391829007 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	1.7	50	56.7	110	70-130	

MATRIX SPIKE SAMPLE:	2303229						
Parameter	Units	70391829008 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	1.4	50	55.4	108	70-130	

SAMPLE DUPLICATE: 2303225

Parameter	Units	70391829007 Result	Dup Result	RPD	Max RPD	Qualifiers
Lead	ug/L	1.7	1.7	1	20	

SAMPLE DUPLICATE: 2303227

Parameter	Units	70391829008 Result	Dup Result	RPD	Max RPD	Qualifiers
Lead	ug/L	1.4	1.4	1	20	

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



QUALITY CONTROL DATA

Project: BURNT HILLS MS 11/6
Pace Project No.: 70391829

QC Batch:	430489	Analysis Method:	EPA 200.8
QC Batch Method:	EPA 200.8	Analysis Description:	200.8 MET No Prep Drinking Water
		Laboratory:	Pace Analytical Services - Melville

Associated Lab Samples: 70391829027, 70391829028, 70391829029, 70391829030, 70391829031, 70391829032, 70391829033

METHOD BLANK: 2303233 Matrix: Water
Associated Lab Samples: 70391829027, 70391829028, 70391829029, 70391829030, 70391829031, 70391829032, 70391829033

Parameter	Units	Blank Result	Reporting Limit	Analyzed	Qualifiers
Lead	ug/L	<1.0	1.0	12/01/25 17:59	

LABORATORY CONTROL SAMPLE: 2303234

Parameter	Units	Spike Conc.	LCS Result	LCS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	50	51.8	104	85-115	

MATRIX SPIKE SAMPLE: 2303236

Parameter	Units	70391829027 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	<1.0	50	58.1	116	70-130	

MATRIX SPIKE SAMPLE: 2303238

Parameter	Units	70391829028 Result	Spike Conc.	MS Result	MS % Rec	% Rec Limits	Qualifiers
Lead	ug/L	14.8	50	67.4	105	70-130	

SAMPLE DUPLICATE: 2303235

Parameter	Units	70391829027 Result	Dup Result	RPD	Max RPD	Qualifiers
Lead	ug/L	<1.0	<1.0		20	

SAMPLE DUPLICATE: 2303237

Parameter	Units	70391829028 Result	Dup Result	RPD	Max RPD	Qualifiers
Lead	ug/L	14.8	15.0	1	20	

Results presented on this page are in the units indicated by the "Units" column except where an alternate unit is presented to the right of the result.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



QUALIFIERS

Project: BURNT HILLS MS 11/6

Pace Project No.: 70391829

DEFINITIONS

DF - Dilution Factor, if reported, represents the factor applied to the reported data due to dilution of the sample aliquot.

ND - Not Detected at or above adjusted reporting limit.

TNTC - Too Numerous To Count

J - Estimated concentration above the adjusted method detection limit and below the adjusted reporting limit.

MDL - Adjusted Method Detection Limit.

PQL - Practical Quantitation Limit.

RL - Reporting Limit - The lowest concentration value that meets project requirements for quantitative data with known precision and bias for a specific analyte in a specific matrix.

S - Surrogate

1,2-Diphenylhydrazine decomposes to and cannot be separated from Azobenzene using Method 8270. The result for each analyte is a combined concentration.

Consistent with EPA guidelines, unrounded data are displayed and have been used to calculate % recovery and RPD values.

LCS(D) - Laboratory Control Sample (Duplicate)

MS(D) - Matrix Spike (Duplicate)

DUP - Sample Duplicate

RPD - Relative Percent Difference

NC - Not Calculable.

SG - Silica Gel - Clean-Up

U - Indicates the compound was analyzed for, but not detected.

N-Nitrosodiphenylamine decomposes and cannot be separated from Diphenylamine using Method 8270. The result reported for each analyte is a combined concentration.

Reported results are not rounded until the final step prior to reporting. Therefore, calculated parameters that are typically reported as "Total" may vary slightly from the sum of the reported component parameters.

Pace Analytical is TNI accredited. Contact your Pace PM for the current list of accredited analytes.

TNI - The NELAC Institute.

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.



QUALITY CONTROL DATA CROSS REFERENCE TABLE

Project: BURNT HILLS MS 11/6

Pace Project No.: 70391829

Lab ID	Sample ID	QC Batch Method	QC Batch	Analytical Method	Analytical Batch
70391829001	MS-002	EPA 200.8	430441		
70391829002	MS-004	EPA 200.8	430441		
70391829003	MS-009	EPA 200.8	430441		
70391829004	MS-010	EPA 200.8	430441		
70391829005	MS-024	EPA 200.8	430441		
70391829006	MS-025	EPA 200.8	430441		
70391829007	MS-058	EPA 200.8	430488		
70391829008	MS-059	EPA 200.8	430488		
70391829009	MS-060	EPA 200.8	430488		
70391829010	MS-061	EPA 200.8	430488		
70391829011	MS-062	EPA 200.8	430488		
70391829012	MS-063	EPA 200.8	430488		
70391829013	MS-072	EPA 200.8	430488		
70391829014	MS-073	EPA 200.8	430488		
70391829015	MS-080	EPA 200.8	430488		
70391829016	MS-081	EPA 200.8	430488		
70391829017	MS-084	EPA 200.8	430488		
70391829018	MS-085	EPA 200.8	430488		
70391829019	MS-086	EPA 200.8	430488		
70391829020	MS-088	EPA 200.8	430488		
70391829021	MS-089	EPA 200.8	430488		
70391829022	MS-104	EPA 200.8	430488		
70391829023	MS-105	EPA 200.8	430488		
70391829024	MS-107	EPA 200.8	430488		
70391829025	MS-151	EPA 200.8	430488		
70391829026	MS-153	EPA 200.8	430488		
70391829027	MS-154	EPA 200.8	430489		
70391829028	MS-155	EPA 200.8	430489		
70391829029	MS-156	EPA 200.8	430489		
70391829030	MS-157	EPA 200.8	430489		
70391829031	MS-158	EPA 200.8	430489		
70391829032	MS-159	EPA 200.8	430489		
70391829033	MS-160	EPA 200.8	430489		

REPORT OF LABORATORY ANALYSIS

This report shall not be reproduced, except in full,
without the written consent of Pace Analytical Services, LLC.

Pace® Location Requested (City/State):
Pace Analytical Long Island NY
575 Broad Hollow Rd, Melville, NY 11747

Company Name:
Intertek-PSI
Street Address:
17 British American Blvd, Latham, NY 12110

Customer Project #:
Project Name:
Burnt Hills Ballston Lake CSD

Site Collection Info/Facility ID (as applicable):

Contact/Report To: William Kotas
Phone #: (518) 377-9841
E-Mail: william.kotas@intertek.com
Cc E-Mail:

Invoice To: PSI Latham Accounts Payable
Invoice E-Mail: LathamAR@intertek.com
Purchase Order # (if applicable):
Quote #: CR-BOCES

Time Zone Collected: [] AK [] PT [] MT [] CT [X] ET

Data Deliverables:

[] Level II [] Level III [] Level IV
[] EQUIS
[] Other

Rush (Pre-approval required):
[] 2 Day [] 3 day [] 5 day [] Other
Date Results Requested: Standard 10 business day

DW PWSID # or WW Permit # as applicable:
Field Filtered (if applicable): [] Yes [] No
Analysis:

* Matrix Codes (Insert in Matrix box below): Drinking Water (DW), Ground Water (GW), Waste Water (WW), Product (P), Soil/Solid (SS), Oil (OL), Wipe (WP), Tissue (TS), Bioassay (B), Vapor (V), Other (OT), Surface Water (SW), Sediment (SED), Sludge (SL), Caulk

Customer Sample ID	Matrix *	Comp / Grab	Collected		Composite End	Res. CL2	Number & Type of Containers	
			Date	Time			Plastic	Glass
MS-002	DW	G	11/6/25	04:47			1	
MS-004				04:49				
MS-009				04:52				
MS-010				04:54				
MS-024				05:01				
MS-025				05:03				
MS-058				05:08				
MS-059				05:08				
MS-060				05:08				

Customer Remarks / Special Conditions / Possible Hazards:

Collected By: William A. Kotas
Signature:

Additional Instructions from Pace®:

Coolers: Thermometer ID: Correction Factor (°C): Obs. Temp (°C) Corrected Temp (°C)

Requisitioned by/Company: Signature: Date/Time: 11/11/25 11:30

Requisitioned by/Company: Signature: Date/Time: 11/11/25 12:30

Requisitioned by/Company: Signature: Date/Time: 11/11/25 07:30

Requisitioned by/Company: Signature: Date/Time: 11/11/25 07:30

Received by/Company: Signature: Date/Time: 11/11/25 11:30

Received by/Company: Signature: Date/Time: 11/11/25 12:30

Received by/Company: Signature: Date/Time: 11/11/25 07:30

Received by/Company: Signature: Date/Time: 11/11/25 07:30

Tracking Number:

Delivered by: [] In-Person [] Courier

[] FedEx [] UPS [] Other

Page: of

ENV-FRM-CORQ-0019_v01_082123 ©

Submitting a sample via this chain of custody constitutes acknowledgment and acceptance of the Pace® Terms and Conditions found at <https://www.pacelabs.com/resource-library/resource/pace-terms-and-conditions/>

LAB USE ONLY- Affix Workorder/Login Label Here

Scan QR Code for instructions

CHAIN-OF-CUSTODY Analytical Request Document

Chain-of-Custody is a LEGAL DOCUMENT- Complete all relevant fields

Company Name: Intertek-PSI
Street Address: 17 British American Blvd, Latham, NY 12110
Customer Project #: Burnt Hills Ballston Lake CSD
Project Name:
Site Collection Info/Facility ID (as applicable):

Contact/Report To: William Kotas
Phone #: (518) 377-9841
E-Mail: william.kotas@intertek.com
Cc E-Mail:
Invoice To: PSI Latham Accounts Payable
Invoice E-Mail: LathamAR@intertek.com
Purchase Order # (if applicable):
Quote #: CR-BOCES

Time Zone Collected: [] AK [] PT [] MT [] CT [] ET
Data Deliverables: [] Level II [] Level III [] Level IV [] EQUIS [] Other

Regulatory Program (DW, RCRA, etc.) as applicable: NY Lead in School DW
Rush (Pre-approval required): [] 1-2 Day [] 3-5 day [] 5-10 day [] Other
Date Results Requested: Standard 10 business day

Field Filtered (if applicable): [] Yes [] No
Analysis: DW PWSID # or WW Permit # as applicable:
Analysis: Field Filtered (if applicable): [] Yes [] No

Matrix Codes (Insert in Matrix box below): Drinking Water (DW), Ground Water (GW), Waste Water (WW), Product (P), Soil/Solid (SS), Oil (OL), Wipe (WP), Tissue (TS), Bioassay (B), Vapor (V), Other (OT), Surface Water (SW), Sediment (SD), Sludge (SL), Caulk

Customer Sample ID: MS-061, MS-062, MS-063, MS-072, MS-073, MS-080, MS-081, MS-084, MS-085

Matrix: DW
Comp / Grab: G
Date: 11/16/25
Time: 05:09, 05:10, 05:10, 05:13, 05:14, 05:17, 05:18, 05:20, 05:25

Collected (or Composite Start) Date: 11/16/25, Time: 05:09, 05:10, 05:10, 05:13, 05:14, 05:17, 05:18, 05:20, 05:25

Res. CL2: [] Plastic [] Glass
Number & Type of Containers: 1

Composite End Date: [] Date: [] Time: []

Collected By: William A. Kotas
Printed Name: William A. Kotas
Signature: [Signature]

Received by/Company: [Signature]
Received by/Company: [Signature]
Received by/Company: [Signature]
Received by/Company: [Signature]

Date/Time: 11/16/25 11:30, 11/14/25 12:30, 11/15/25 08:17, 11/15/25 12:05

Relinquished by/Company: [Signature]
Relinquished by/Company: [Signature]
Relinquished by/Company: [Signature]
Relinquished by/Company: [Signature]

Date/Time: 11/14/25 11:30, 11/15/25 08:30, 11/15/25 08:37, 11/15/25 12:05

Delivered by: [] In-Person [] Courier [] FedEx [] UPS [] Other

Page: 1 of 1

[illegible]

Client:

Inter-latham
Burnt Hills MS

Work ID:

Profile #:

10367

COC Page

of

☐ Use Point Number Spreadsheet

☐ Multiday Project

☐ Add SCLOGFD to first sample for field charge

Matrix	1	2	3	4	5	6	7	8	9	10	11	12
V'G9U												
V'G8C												
V'G9H												
V'G8S												
V'G9T												
D'G9Y												
D'G9P												
D'G9A												
D'G6T												
D'G9S												
A'G4E												
A'G3T												
A'G2R												
A'G1T												
A'G1H												
A'G1A												
A'G5U												
A'G44												
CG1U												
WG90												
WG40												
BP4U												
BP3U												
BP2U												
BP1U												
BP3S												
BP2S												
BP4N												
BP3N												
BP2N												
BP3C												
BP3T												
BP3S												
BP3R												
BP1Z												
BP1N												
BP1B												
SP5T												
R												
WG2U												
WGFU												
WGKU												
WGDU												
ZPLC												
GN												
WP												
TEDL												
B'G1H												
LIHG												
B'G1N												
IOC												
SOC												

WO#: 70391829

Client Name:

Finkler - Latham

Project #

PM: LAB

Due Date: 12/03/25

CLIENT: INTER-LATHAM

Courier: ☐ Fed Ex ☐ UPS ☐ USPS ☐ Client ☐ Commercial ☒ Pace ☐ Other

Tracking #:

Custody Seal on Cooler/Box Present: ☐ Yes ☒ No Seals intact: ☐ Yes ☒ No Temperature Blank Present: ☐ Yes ☒ No

Packing Material: ☐ Bubble Wrap ☐ Bubble Bags ☒ Ziploc ☐ None ☐ Other Type of Ice: ☒ Wet ☐ Blue ☐ None

Thermometer Used: TH211 Correction Factor: 0.1 ☒ Samples on ice, cooling process has begun

Cooler Temperature (°C): 17.2 Cooler Temperature Corrected (°C): 17.3 Date/Time 5035A kits placed in freezer

Temp should be above freezing to 6.0°C

USDA Regulated Soil (☐ N/A, water sample)

Did samples originate in a quarantine zone within the United States: AL, AR, CA, FL, GA, ID, LA, MS, NC, NM, NY, OK, OR, SC, TN, TX, or VA (check map)? ☐ Yes ☒ No

Did samples originate from a foreign source including Hawaii and Puerto Rico? ☐ Yes ☒ No

If Yes to either question, fill out a Regulated Soil Checklist (ENV-FRM-MELV-0076) and include with SCUR/COC paperwork.

Date and Initials of person examining contents:

COMMENTS:	
Chain of Custody Present: ☒ Yes ☐ No	1.
Chain of Custody Filled Out: ☒ Yes ☐ No	2.
Chain of Custody Relinquished: ☒ Yes ☐ No	3.
Sampler Name & Signature on COC: ☒ Yes ☐ No ☐ N/A	4.
Samples Arrived within Hold Time: ☒ Yes ☐ No	5.
Short Hold Time Analysis (<72hr): ☐ Yes ☒ No	6.
Rush Turn Around Time Requested: ☐ Yes ☒ No	7.
Sufficient Volume: (Triple volume provided for MS/MSD) ☒ Yes ☐ No	8.
Correct Containers Used: ☒ Yes ☐ No	9.
-Pace Containers Used: ☒ Yes ☐ No	10.
Containers Intact: ☒ Yes ☐ No	11.
Filtered volume received for Dissolved tests: ☐ Yes ☐ No ☐ N/A	12.
Sample Labels match COC: ☒ Yes ☐ No	
-Includes date/time/ID/Analysis Matrix: <u>SL</u> <u>WT</u> <u>OIL</u> <u>OTHER</u>	

Date and Initials of person checking preservation:

All containers needing preservation have been pH paper Lot # <u>210525</u> ☒ Yes ☐ No ☐ N/A	13. ☐ HNO ₃ ☐ H ₂ SO ₄ ☐ NaOH ☐ HCl
All containers needing preservation are found to be in compliance with method recommendation? (HNO ₃ , H ₂ SO ₄ , HCl, NaOH > 9 Sulfide, ☒ Yes ☐ No ☐ N/A NAOH > 12 Cyanide) Exceptions: VOA, Coliform, TOC/DOC, Oil and Grease, DRO/8015 (water). Per Method, VOA pH is checked after analysis	Sample #
Samples checked for dechlorination: ☐ Yes ☐ No ☐ N/A	Initial when completed: Lot # of added preservative: Date/Time preservative added:
KI starch test strips Lot #	14. Positive for Res. Chlorine? ☐ Y ☒ N
Residual chlorine strips Lot #	15. Positive for Sulfide? ☐ Y ☒ N
SM 4500 CN samples checked for sulf ☐ Yes ☐ No ☐ N/A	16.
Lead Acetate Strips Lot #	17.
Headspace in ALK Bottle (>6mm): ☐ Yes ☐ No ☐ N/A	
Headspace in VOA Vials (>6mm): ☐ Yes ☐ No ☐ N/A	
Trip Blank Present: ☐ Yes ☐ No ☐ N/A	
Trip Blank Custody Seals Present: ☐ Yes ☐ No ☐ N/A	

Client Notification/ Resolution:

Field Data Required? Y / N

Person Contacted:

Date/Time:

Comments/ Resolution:

* PM (Project Manager) review (which includes the SCUR) is documented electronically in LIMS.